

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 10 2008

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130292</u> ✓
Name <u>Gary Lowe</u>		ID# <u>7219</u> Date <u>12/3/08</u> ✓
A	Agency <u>Haines P.D.</u>	Phone # _____
Instrument Location <u>Haines P.D.</u> ✓		
Alco S/N <u>X124631</u> ✓ Target Value <u>.082</u> ✓ High Pressure <u>350</u> psi		
B	Alco Test Values <u>.084</u> ✓ <u>.084</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓		<u>CS</u> <u>12/3/08</u>
(OVER)		


Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary A Lowe, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/23/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130292 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

DECEMBER 03, 2008 ✓ *DOB 12/23/08*

OPERATOR'S NAME:
LOWE/GARY/A
OPERATOR'S NUMBER: 7219
SUBJECT'S LAST NAME:

VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 12345
DEPT/AGENCY: HNS1
CASE/REPORT: 00-120308
TEST TYPE: V
ALCO TARGET VALUE: .002 ✓
ALCO S/N: X124631 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.63 in		
ALCO TARGET	.003	15:42
BLANK TEST	.000	15:43
INTERNAL STANDARD	VERIFIED	15:43
ALCO TV 30.63 in	.004 ✓	15:44
BLANK TEST	.000	15:44
SUBJECT SAMPLE	.000	15:45
BLANK TEST	.000	15:46
ALCO TV 30.63 in	.004 ✓	15:46
BLANK TEST	.000	15:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130292

DECEMBER 03, 2008 ✓
TIME 15:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 30.64 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~